National Bureau of Standards
Library, E-01 Admin. Bldg.
OCT 6 1981
191130
QC
100
.U5753

NBS TECHNICAL NOTE 935

U.S. DEPARTMENT OF COMMERCE / National Bureau of Standards

Measurement of Energy Irradiance from Single Pulse Sources

QC
100
U5753
No.935
1977
c.2

NATIONAL BUREAU OF STANDARDS

The National Bureau of Standards [1] was established by an act of Congress March 3, 1901. The Bureau's overall goal is to strengthen and advance the Nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research and provides: (1) a basis for the Nation's physical measurement system, (2) scientific and technological services for industry and government, (3) a technical basis for equity in trade, and (4) technical services to promote public safety. The Bureau consists of the Institute for Basic Standards, the Institute for Materials Research, the Institute for Applied Technology, the Institute for Computer Sciences and Technology, and the Office for Information Programs.

THE INSTITUTE FOR BASIC STANDARDS provides the central basis within the United States of a complete and consistent system of physical measurement; coordinates that system with measurement systems of other nations; and furnishes essential services leading to accurate and uniform physical measurements throughout the Nation's scientific community, industry, and commerce. The Institute consists of the Office of Measurement Services, the Office of Radiation Measurement and the following Center and divisions:

Applied Mathematics — Electricity — Mechanics — Heat — Optical Physics — Center for Radiation Research: Nuclear Sciences; Applied Radiation — Laboratory Astrophysics [2] — Cryogenics [2] — Electromagnetics [2] — Time and Frequency [2].

THE INSTITUTE FOR MATERIALS RESEARCH conducts materials research leading to improved methods of measurement, standards, and data on the properties of well-characterized materials needed by industry, commerce, educational institutions, and Government; provides advisory and research services to other Government agencies; and develops, produces, and distributes standard reference materials. The Institute consists of the Office of Standard Reference Materials, the Office of Air and Water Measurement, and the following divisions:

Analytical Chemistry — Polymers — Metallurgy — Inorganic Materials — Reactor Radiation — Physical Chemistry.

THE INSTITUTE FOR APPLIED TECHNOLOGY provides technical services to promote the use of available technology and to facilitate technological innovation in industry and Government; cooperates with public and private organizations leading to the development of technological standards (including mandatory safety standards), codes and methods of test; and provides technical advice and services to Government agencies upon request. The Institute consists of the following divisions and Centers:

Standards Application and Analysis — Electronic Technology — Center for Consumer Product Technology: Product Systems Analysis; Product Engineering — Center for Building Technology: Structures, Materials, and Life Safety; Building Environment; Technical Evaluation and Application — Center for Fire Research: Fire Science; Fire Safety Engineering.

THE INSTITUTE FOR COMPUTER SCIENCES AND TECHNOLOGY conducts research and provides technical services designed to aid Government agencies in improving cost effectiveness in the conduct of their programs through the selection, acquisition, and effective utilization of automatic data processing equipment; and serves as the principal focus within the executive branch for the development of Federal standards for automatic data processing equipment, techniques, and computer languages. The Institute consists of the following divisions:

Computer Services — Systems and Software — Computer Systems Engineering — Information Technology.

THE OFFICE FOR INFORMATION PROGRAMS promotes optimum dissemination and accessibility of scientific information generated within NBS and other agencies of the Federal Government; promotes the development of the National Standard Reference Data System and a system of information analysis centers dealing with the broader aspects of the National Measurement System; provides appropriate services to ensure that the NBS staff has optimum accessibility to the scientific information of the world. The Office consists of the following organizational units:

Office of Standard Reference Data — Office of Information Activities — Office of Technical Publications — Library — Office of International Relations — Office of International Standards.

---

[1] Headquarters and Laboratories at Gaithersburg, Maryland, unless otherwise noted; mailing address Washington, D.C. 20234.

[2] Located at Boulder, Colorado 80302.

al Bureau of Standards

MAR 10 1977

+ acc.

C

00

5753

935

177

.2

Measurement of Energy Irradiance from Single Pulse Sources

t technical note, no. 935

A. R. Schaefer and E. F. Zalewski

Optical Physics Division
Institute for Basic Standards
National Bureau of Standards
Washington, D.C. 20234

U.S. DEPARTMENT OF COMMERCE, Juanita M. Kreps, Secretary

Dr. Betsy Ancker-Johnson, Assistant Secretary for Science and Technology

NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Acting Director

Issued February 1977

National Bureau of Standards Technical Note 935
Nat. Bur. Stand. (U.S.), Tech. Note 935, 32 pages (Feb. 1977)
CODEN: NBTNAE

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1977

For sale by the Superintendent of Documents, U.S. Government Printing Office
Washington, D.C. 20402 - Price 75 cents

Stock No. 003-003-01741-8

There is a minimum charge of $1.00 for each mail order

TABLE OF CONTENTS

PAGE

1. INTRODUCTION 1

2. THEORY OF THE MEASUREMENT 2

3. DETAILED EXPERIMENTAL CONSIDERATIONS 3

   3.1 The Shutter Mechanism and the Evaluation of the Time Integral 3

   3.2 Errors Inherent in a Photopic Measurement 5

   3.3 Electronic Considerations 6

4. MEASUREMENT PROCEDURE 7

5. DATA ANALYSIS 9

   5.1 A Mathematical Model of the Illuminance from an Apertured Cylindrical Lamp 9

   5.2 Treatment of the Data 12

6. CURVE-FITTING 17

7. SUMMARY 21

8. REFERENCES 21

LIST OF FIGURES

1. Shuttered time response of detector illustrating integration of total pulse. (A) Detector has same shape as sector disk. (B) Circular detector 4

2. Experimental setup for measurement 8

3. Illustration of model for analysis of variation of illuminance as a function of detector distance from flashlamp 10

4. Plot of illuminance-distance squared product versus distance of a given flashlamp with detector 1 13

5. Plot of illuminance-distance squared product versus distance of a given flashlamp with detector 2 14

6. Data of Fig. 4 with theoretical curve derived from Eq. 25 . . . 15

7. Data of Fig. 5 with theoretical curve derived from Eq. 25 . . . 16

8. Data of Fig. 4 with curve fitting of Eq. 17 18

PAGE

9. Data of Fig. 5 with curve fitting of Eq. 17 19

10. Data of Fig. 5 with three curve fitting equations. __.__. is Eq. 27. __ __ __ is Eq. 28. _____ is Eq. 26 20

LIST OF TABLES

1. Illuminance of a Given Flashlamp as a Function of Distance as Measured by Detector 1 22

2. Illuminance of a Given Flashlamp as a Function of Distance as Measured by Detector 2 24

MEASUREMENT OF ENERGY IRRADIANCE FROM SINGLE PULSE SOURCES

A. R. Schaefer and E. F. Zalewski

A method of measuring the energy irradiance from a single pulsed source, such as a xenon flashtube, is presented. Details of the approach used in this particular measurement are given, along with a discussion of the data analysis. In particular, we consider some of the inherent problems encountered in curve fitting which are also common to other radiometric and photometric measurements.

Key Words: Curve fitting; energy irradiance; flashtube, pulsed source; radiometry.

****

1. Introduction

In this paper we describe a method for measuring the amount of energy per unit area, or energy irradiance [1][1], propagating from a pulsed source such as a xenon flashtube at a given distance from the source. The basic measurement procedure is applicable when the energy is either to be measured spectrally (per unit wavelength) or to be integrated over a spectral bandwidth, such as in the case of photopic (luminous) energy. The latter measurement was performed as a special calibration by NBS. While preparing for this calibration, it was realized that little previous documentation existed concerning this type of measurement. Since there is an apparent need for such measurements, this paper is being written to document the procedure we have used at NBS.

We will present a specific method of calibrating the energy per unit area of a pulsed source using a continuous (non-pulsed) standard of either irradiance or illuminance (luminous intensity). The theoretical approach will be outlined, followed by a detailed discussion of the factors contributing to the uncertainty of the measurement. As a specific example, we will describe the particular measurement performed at NBS. Finally, we will conclude with a discussion of the treatment of the data, the curve fitting and its inherent problems.

2. Theory of the Measurement

In measuring the energy irradiance output of a pulsed source, the first problem to arise comes from the fact that the only basic reference standards currently available are constant power sources, such as irradiance or illuminance standards, the latter being derived to measure the radiant power incident on a given area within a known time interval. The integral of power over this time interval is the total energy. That is,

$$W = \int_{t_1}^{t_2} E(t)\, dt \qquad (1)$$

where E(t) is the power irradiance as a function of time and W is the energy per unit area. A continuous source can be chopped by a rotating blade to provide a single pulse within a measurable time interval and the integration can be accomplished electrically. For our measurements, a silicon photodiode and operational amplifier were used. The emitted current, i(t), was proportional to the irradiance incident upon the detector:

$$i(t) = k_1\, E(t) \qquad (2a)$$

[1]Figures in brackets indicate the literature references at the end of this paper.

If we write Eq. (2a) in terms of the charge Q, then

$$\frac{dQ}{dt} = k_1 E(t) \quad . \tag{2b}$$

We used an electrometer in its charge measuring mode, in which it behaves essentially like a capacitor. It produces a voltage, V, proportional to the charge, so that,

$$V = \frac{Q}{C} \quad . \tag{3a}$$

Since Q and V change in time we can differentiate and obtain

$$\frac{dV}{dt} = \frac{1}{C} \frac{dQ}{dt} \quad . \tag{3b}$$

Substituting into Eq. (3b) yields

$$C \frac{dV}{dt} = k_1 E(t) \tag{4a}$$

or

$$\frac{dV}{dt} = k E(t) \quad . \tag{4b}$$

Therefore, over a time interval t_1 to t_2:

$$V = k \int_{t_1}^{t_2} E(t) \, dt \quad . \tag{5}$$

Equation (5) is the working relation of the measurement system: a power irradiance pulse is received by the detector, and subsequently a voltage, proportional to the total energy in the pulse, is read on the electrometer capacitor. The calibration is made absolute by the use of a reference standard in the determination of k.

Determination of the system calibration constant, k, can be accomplished by measuring the voltage V_0 that results after applying a pulse of known power irradiance. The energy irradiance integral can be evaluated in the following manner. First, a known constant power irradiance, E_0, is produced at the detector by a suitable standard incandescent lamp. A shutter is placed between the standard lamp and the detector. If one assumes that the shutter instantaneously opens at time t_1 and closes at t_2, then a pulse of known power irradiance is incident on the detector in the specified time interval. From Eq. (5) we can obtain the value of k:

$$k = \frac{V_0}{\int_{t_1}^{t_2} E_0 \, dt} = \frac{V_0}{E_0(t_2 - t_1)} \tag{6}$$

3. Detailed Experimental Considerations

We have outlined the essentials of the method whereby energy irradiance calibrations can be made. We shall now go into several more detailed, practical aspects of the measurement which, if not given due attention, could have an adverse effect upon the results.

3.1 The Shutter Mechanism and the Evaluation of the Time Integral

The shutter that provides the time calibration information must have the following characteristics. It must, of course, provide a stable, repeatable, and well-determined open time. In addition, it should expose the entire detector surface to the source in a well defined manner. In this regard, one must be wary of accidentally using the shutter in such a way that the detector will not be illuminated in the same manner by both the shutter and standard lamp combination and the pulsed source to be calibrated. This would be the case, for instance, if one used a shutter opening that was not larger than the detector aperture.

The shutter used for this calibration was a single opening sector disc of 63°, rotating once every half second. The circular detector, area about 1 cm^2, was placed on a radius about 15 cm from the axis of the sector disc. Under these conditions the shutter opening or closing can be approximated by a knife edge shadow moving across the detector in a period about one sixteenth that of the total open time. Let us assume for the purpose of discussion that we have a detector that has the same shape as the opening in the sector disc plus uniform and linear responsivity. The time dependent output of the detector will then be a trapezoid, as outlined in Fig. 1a. Segments t_1 to t_3 and t_4 to t_6 will be straight lines, since the area irradiated is changing linearly with time. In order to apply Eq. (6) one must know the area of this pulse from t_1 to t_6 assuming height E_0. The area of the trapezoid is given by $A = 1/2(t_4 - t_3 + t_6 - t_1)\, E_0$. Since

$$t_3 - t_2 = t_2 - t_1 = t_6 - t_5 = t_5 - t_4 \qquad (8)$$

then

$$t_4 - t_3 = t_5 - t_2 - 2(t_3 - t_2) \qquad (9)$$

and

$$t_6 - t_1 = t_5 - t_2 + 2(t_2 - t_1) \quad . \qquad (10)$$

Therefore, the area of the trapezoid is simply

$$A = (t_5 - t_2)\, E_0 \quad . \qquad (11)$$

That is, the area of a trapezoidal pulse is equal to the area of a rectangular pulse drawn through half-power points t_2 and t_5 and of height E_0. This is a useful operational simplification.

In reality, the detector used is generally not rectangular in shape, but circular. Hence the segments t_1 to t_3 and t_4 to t_6 are not straight lines, since the detector area irradiated will not be a linear function of time. They will have a shape similar to that shown in Fig. 1b.

One can calculate from the area of a circular segment, what the area of this pulse shape will be. Let





Fig. 1: Shuttered time response of detector illustrating integration of total pulse. A) Detector has same shape as sector disk. B) Circular detector.

$$I_1 = \frac{E_0}{2} - \left[x \sqrt{(E_0/\pi) - x^2} + (E_0/\pi) \sin^{-1}(x \sqrt{\pi/E_0}) \right] , \quad (12a)$$

$$I_2 = \frac{E_0}{2} + \left[t \sqrt{(E_0/\pi) - t^2} + (E_0/\pi) \sin^{-1}(t \sqrt{\pi/E_0}) \right] , \quad (12b)$$

$$I_3 = \frac{E_0}{2} + \left[x \sqrt{(E_0/\pi) - x^2} + (E_0/\pi) \sin^{-1}(x \sqrt{\pi/E_0}) \right] , \quad (12c)$$

$$I_4 = \frac{E_0}{2} - \left[t \sqrt{(E_0/\pi) - t^2} + (E_0/\pi) \sin^{-1}(t \sqrt{\pi/E_0}) \right] , \quad (12d)$$

where

$$x = \sqrt{E_0/\pi} - t \quad ; \quad (12e)$$

then the area is

$$A = \int_{t_1}^{t_2} I_1 dt + \int_{t_2}^{t_3} I_2 dt + \int_{t_3}^{t_4} E_0 dt + \int_{t_4}^{t_5} I_3 dt + \int_{t_5}^{t_6} I_4 dt \quad . \quad (13)$$

Working through Eq. (13), and using the fact that, from symmetry, Eqs. 8, 9, and 10 are again true, one gets the result

$$A = E_0(t_5 - t_2) \quad (14)$$

This is the same result obtained for the simpler case treated earlier.

From the foregoing discussion we see that in practice one utilizes Eq. 6 with the following modification: instead of $(t_2 - t_1)$, one uses the time interval between half power points, $(t_5 - t_2)$.

We should emphasize the importance of maintaining a constant measurable shutter motion. Due to mechanical imperfections in the system, one should not assume that an ordinary synchronous motor will drive a rotary shutter wheel at the precisely expected rate. The importance of the accuracy of the time measurement to the overall calibration is obvious.

3.2 Errors Inherent in a Photopic Measurement

There are specific problems that arise from doing this calibration photometrically. In this case, the quantity of interest is the photopic[2] energy irradiance obtained from the time integral of the photopic power irradiance (illuminance) rather than the spectral power irradiance. The measurement is carried out exactly as described above, except that a photopic detector must be used. It must be calibrated with a photometric standard, such as luminous intensity. The uncertainty arises in answer to the question: How well does the detector response actually correspond to the CIE defined $V(\lambda)$ curve?

The photometric standard that was used to calibrate the detector was an incandescent source having a spectral output resembling that of a thermal radiator at 2856 K. This calibrated detector was then used to measure the output of xenon flashtubes which have a very different spectrum. They have a much higher relative energy in the blue than the incandescent standards. These spectral differences coupled with the detector response deviations from the $V(\lambda)$ curve can lead to a sizeable measurement error.

In order to approximate the probable magnitude of this error, hypothetical spectral response curves of several photopic detectors* were used in calculating a theoretical calibration between a blackbody incandescent source at 2856 K and the spectral output of a typical xenon flashlamp. We used the relation:

$$E_{ij} = K_m \int_{380\ nm}^{760\ nm} D_i(\lambda)\ E_j(\lambda)\ d\lambda \qquad (15)$$

where E_{ij} represents the illuminance measured by a detector i from a source j, $E_j(\lambda)$ is the spectral irradiance from the j^{th} source, $D_i(\lambda)$ is the spectral responsivity of the i^{th} detector, and K_m is the maximum luminous efficacy. The value of K_m need not be specified since it cancels out in the comparison. For the perfect detector, $D_i(\lambda)$ is the CIE $V(\lambda)$ curve. The integrals were evaluated numerically using Simpson's Rule[3] with 10 nm wavelength intervals.

The hypothetical detector responses usually deviated from the defined $V(\lambda)$ curve by one or two percent, and sometimes more in the extreme red and blue regions. The results of these calculated calibrations differed from those obtained using the perfect photopic detector by as much as 5%. This was in agreement with the results of actual experimental calibrations. The same flashlamp and standard lamp were compared under identical conditions with two different types of photopic detectors (silicon photodiodes from different manufacturers). This experiment yielded results which disagreed by about 4%. The experimental precision of the measurement in each case was about ±0.15%. It is obvious that accurate photometric measurements of this type require precise characterization of the spectral response of the detector being used.

3.3 Electronic Considerations

The following experimental considerations hold true for many types of measurements in addition to the one under discussion. It is important not to overlook them so they will be briefly mentioned at this time.

First, the detector and amplifier should have a response time much faster than the rise times of the pulses being measured. This is to insure that the calibration pulse from the shuttered lamp and the flash pulse are both integrated properly in spite of their different waveforms. This can easily be checked with a variable high speed sector disc and an oscilloscope to monitor the detector-amplifier output as a function of pulse frequency. For this calibration, an FET operational amplifier with low capacitance in the feedback circuit was used. This device proved to have an adequately short rise time. It is the second version of the dc amplifier described in Technical Note 594-2.[4]

Another problem is detector linearity. Although one can attempt to set the energy irradiance from the shuttered source approximately equal to that of the pulse, this is difficult to do exactly. In addition, energy irradiance measurements may have to be made at a number of different distances, and it is impractical to recalibrate the detector for each distance from the flashlamp. Finally, the detector-amplifier obviously must respond to a large variation in signal level during each pulse. It is, therefore, very important to ascertain that the detector and amplifier respond linearly. For the detectors used in these measurements the linearity was checked using the inverse square approximation with a luminous intensity standard lamp. They appeared to be acceptably linear; however, as we will show in Section 5.2, this was not an entirely satisfactory test.

Additional details which must be considered are the accurate measurement of the detector to source distance, the repeatability of the positioning of the lamp and detector, standard lamp accuracy, flashlamp repeatability, noise pickup, and detector-amplifier stability.

---

*Information obtained from the specification sheets of several manufacturers.

4. Measurement Procedure

The calibration of a flashlamp (with power supply included) was a two part operation: calibration of the detector-amplifier system for photopic energy using a luminous intensity standard lamp with a timing shutter, and subsequent calibration of the flashlamp using the calibrated detector. It was found experimentally that the system drift during the course of a flashlamp measurement was negligible, hence the calibrations of the detector-amplifier were actually done last, after a given flashlamp measurement. The flashlamp was placed in front of the standard lamp as shown in Fig. 2. In this way, the flashlamp could be removed after measurement without disturbing the detector and standard lamp arrangement.

After aligning the optical bench, detector, and standard lamp with a He-Ne laser to define the optic axis, the flashlamp was installed in the position shown. The flashlamp assembly consisted of a thin vertical discharge tube behind a rectangular aperture and an absorbing glass filter. Using the laser, the flashlamp assembly was centered on the optic axis and aligned by autocollimating the laser reflection from the filter glass. An end gauge was used to measure the separation distances from the detector to the flashlamp aperture. The detector was then placed at the smallest source-detector separation distance to be measured in the experiment. Several flash readings were observed on the oscilloscope to select the proper amplifier and electrometer gains, allowing maximum readings without overload distortion. These settings were unchanged for the remainder of the calibration of that flashlamp and for the subsequent detector calibration using the incandescent lamp. Several flashes were first fired on a "cold" tube to allow its output to stabilize before recording readings. Prior to making a reading, the flashtube power supply was allowed ample time to recharge after a previous flash. A gate switch was then opened to the electrometer, and bucking current from the amplifier was set to null the electrometer output reading at zero. Then the lamp was flashed via a remote trigger switch, and the electrometer gate closed, "locking in" a voltage reading on the electrometer capacitor. Several measurements were made at each distance position.

When these measurements were completed, the detector was left at the maximum distance point and the flashlamp assembly removed as noted earlier. The shutter was installed in front of the detector and the standard lamp turned on and warmed up. The shutter was then started and its time interval determined with an oscilloscope. The oscilloscope time base was calibrated with standard frequency signals. After the lamp warmed up, the detector was momentarily shuttered in order to zero the background current on the detector amplifier. Then a series of readings were made as follows. The pulses were spaced at least 1/2 second apart so they would be easy to isolate with the electrometer gate switch. When the pulses were exactly the proper time width, the gate was momentarily closed, then opened, isolating one pulse. The reading was recorded and the electrometer cleared for the next sample. These readings were then averaged to yield a value for k using Eq. (6). As mentioned earlier, the calibration lamp was a standard of luminous intensity. One must therefore assume that the inverse square approximation is valid. Hence the form of Eq. (6) actually used was

$$k = \frac{V_0(D + \delta)^2}{I_0(t_5 - t_2)} \quad . \tag{16}$$

Here I_0 is the luminous intensity of the standard lamp, D is the measured distance, δ is the correction to the "light center" distance and $(t_5 - t_2)$ is the time interval at the pulse half-height. The value of δ was determined separately by fitting* the measured distances D_i (from about 1.5 to 3 meters) and amplifier readings R_i (assumed proportional to the illuminance) to the following equation

$$R_i\delta^2 + 2R_iD_i\delta + R_iD_i^2 - I_0 = 0 \tag{17}$$

*Some of the difficulties encountered in curve fitting will be taken up in the next Section.



Fig. 2: Experimental setup for measurement. I. Measurement of flashlamp. A) Position of shutter (not used in this part of measurement). B) Detector. C) Flashlamp assembly. D) Lamp standard (in position but not in use in this part of measurement). E) Detector. F) Shutter. G) Position of flashlamp (now removed for this part of measurement). H) Standard lamp.

Finally the flashlamp photopic energy irradiance values, W, were calculated from the measured voltages using the relation.

$$W = V/k \quad . \tag{18}$$

5. Data Analysis

Once the illuminance versus distance data have been obtained, it is often desirable to fit a smooth curve through these points in order to allow interpolation or possibly extrapolation to other desired points. One may be tempted to use some form of the inverse square approximation again, such as

$$I = ED^2 \quad . \tag{19}$$

However, this is not always the best function to describe the variation of illuminance with distance. In fact, by simply assuming there is always a characteristic constant of the source, i.e. the intensity, one can easily be mislead in attempting to interpret the data. Let us first derive the functional form of the relationship that describes the variation of illuminance from the flashlamp as a function of distance.

5.1 A Mathematical Model of the Illuminance from an Apertured Cylindrical Lamp

The xenon flashtubes we measured consisted of a long thin cylinder masked by a rectangular limiting aperture, a fairly common configuration for such sources. In order to make the model of this lamp analytically tractable, we shall assume the following. Consider the narrow cylindrical tube to be a one dimensional, infinitely long vertical linear source of uniform luminance L per unit length. Let us take this to be the x direction. Also consider the source as apertured by an infinitely long horizontal slit of width 2w, in a plane parallel to that of the source and a distance d from it. The axis of the slit is then in the y direction. Finally, consider a rectangular detector, of dimensions 2a in the y' direction and 2b in the x' direction, also in a plane parallel to that of the source and aperture. The detector is at a distance D from the aperture. Figure 3 illustrates this model.

We know that the radiant flux, $d\phi_\alpha$, reaching an element $\alpha(x',y')$ of area dx' dy' of the detector surface from an element of source dx is related to the radiance in the following manner,

$$d\phi_\alpha = \frac{L\, dx \cos\theta\, dx'\, dy'\, \cos\gamma}{r^2} \quad . \tag{20}$$

Here θ and γ are the angles between the distance r and the normals to the source axis and detector plane, respectively. At this point we shall make another approximation. Let us assume that $\theta=\gamma$, which is equivalent to saying that the detector is slightly curved with a radius of curvature equal to D + d, the center of curvature being coincident with the source axis. We can now calculate the illuminance at the point α on the detector which comes from the entire source:

$$E_\alpha = \frac{d\phi_\alpha}{dx'\, dy'} = L \int_{x_{min}}^{x_{max}} \frac{\cos^2\theta}{r^2}\, dx$$

$$= L(D + d)^2 \int_{x_{min}}^{x_{max}} \frac{dx}{[(D + d)^2 + x'^2 + (x - y')^2]^2} \tag{21}$$



Fig. 3: Illustration of model for analysis of variation of illuminance as a function of detector distance from flashlamp.

The limits of this integral are a function of the slit width and can be determined from the geometry of Fig. 3. They are

$$x_{min} = \frac{-w(D + d)}{D}$$

and

$$x_{max} = \frac{w(D + d)}{D}$$

The result of integrating Eq. 21, is then:

$$\frac{d\phi_\alpha}{dx'\,dy'} = \frac{(D+d)^3\,L}{2[(D+d)^2 + x'^2]} \left\{ - \frac{(y'-w)\,D}{[(D+d)^2 + x'^2]\,D^2 + (y'-w)^2\,(D+d)^2} + \frac{(y'+w)\,D}{[(D+d)^2 + x'^2]\,D^2 + (y'+w)^2\,(D+d)^2} + \frac{1}{2[(D+d)^2 + x'^2]^{3/2}} \times \left[\tan^{-1}\left(\frac{-(y'-w)\,(D+d)}{D[(D+d)^2 + x'^2]^{1/2}} \right) - \tan^{-1}\left(\frac{-(y'+w)\,(D+d)}{D[(D+d)^2 + x'^2]^{1/2}} \right) \right] \right\} \qquad (22)$$

To obtain the average illuminance over the entire detector, we must integrate this expression over x' and y' and divide by the detector area. If Eq. (22) is integrated first over y', many terms cancel, leaving

$$\frac{d\phi}{dx'} = \frac{L(D+d)^2}{[(D+d)^2 + x'^2]^{3/2}} \left\{ (a-w)\,\tan^{-1}\left(\frac{-(a-w)\,(D+d)}{D[(D+d)^2 + x'^2]^{1/2}} \right) - (a+w)\,\tan^{-1}\left(\frac{-(a+w)\,(D+d)}{D[(D+d)^2 + x'^2]^{1/2}} \right) \right\} . \qquad (23)$$

At this point we shall make another approximation to aid in carrying out the analysis. Since D is usually large compared with the detector dimensions,

$$(a \pm w)\,(D + d) \ll D[(D + d)^2 + x'^2]^{1/2} \quad .$$

Hence we can replace the arctan terms with the first two terms in the series expansion:

$$\tan^{-1} x \tilde{=} x - \frac{x^3}{3} \quad .$$

We can now carry out the final integration over x' to give the total flux ϕ intercepted by the detector. After some rather tedious algebra and using the arctan approximation just mentioned, we can write the average illuminance, the flux divided by 4ab, in the following form:

$$\bar{E} = \frac{2Lw}{D(D+d)} \left\{ \frac{1}{2} + \frac{(D+d)^2}{2[(D+d)^2 + b^2]} - \frac{b^2}{6(D+d)^2} - \frac{a^2 + w^2}{4D^2} \right.$$

$$\left. - \frac{(a^2+w^2)\ (D+d)^2}{4D^2[(D+d)^2 + b^2]} - \frac{(a^2+w^2)\ (D+d)^4}{6D^2[(D+d)^2 + b^2]^2} + \frac{(a^2+w^2)\ b^2}{12D^2(D+d)^2} \right\} . \qquad (24)$$

We can further simplify Eq. (24) by rewriting it as

$$\bar{E} = \frac{2Lw}{D(D+d)} \left\{ \frac{1}{2} + \frac{1}{2\left[1 + \frac{b^2}{(D+d)^2}\right]} - \frac{b^2}{6(D+d)^2} - \frac{a^2+w^2}{4D^2} \right.$$

$$\left. - \frac{a^2+w^2}{4D^2\left[1 + \frac{b^2}{(D+d)^2}\right]} - \frac{a^2+w^2}{6D^2\left[1 + \frac{b^2}{(D+d)^2}\right]^2} + \frac{(a^2+w^2)\ b^2}{12D^2(D+d)^2} \right\} . \qquad (24a)$$

Then, by dropping terms of order greater than

$$\frac{b^2}{(D + d)^2} << 1$$

the average illuminance is approximately given by

$$\bar{E} = \frac{2Lw}{D(D+d)} \left\{ 1 - \frac{2b^2}{3(D+d)^2} - \frac{2(a^2+w^2)}{3D^2} \right\} \qquad (25)$$

This equation, rather than the inverse square approximation, is a much better model of the relationship between distance and illuminance of the flashlamp sources we studied.

5.2 Treatment of the Data

In Tables 1 and 2 we have listed, and in Figs. 4 and 5 we have plotted, some measurements made on the same flashlamp using two different detectors. Both detectors were photopically corrected silicon photodiodes operated in the photovoltaic mode (zero reverse bias plus a low input impedance amplifier). The detectors were from two different suppliers and, therefore, would be expected to have different characteristics. Instead of plotting just the illuminance vs. distance for these two sets of measurements, we have plotted the product of illuminance and distance squared versus distance. The distance in these data refers to the measured distance from the aperture of the lamp housing to the detector aperture.

One would expect from the inverse square approximation that at large distances the quantity ED^2 would approach a constant value. On this basis we would be inclined to reject the data from detector 2 (Fig. 5). However, let us see how well each of these data sets will fit the equation we have derived in the previous section. For the flashlamp we measured, d was about 4 cm and w was about 0.5 cm. Both detectors were of the same size with a = b $\approx$ 0.5 cm. The only unknown in Eq. (25) is the luminance, L, which can be obtained by using the value of the illuminance measured at one distance. If we arbitrarily choose the points at 1.52 m, the average value of the illuminance measured by detector 1 is 0.281 lux and from detector 2 it is 0.286 lux. The luminance can now be computed and in turn we can compute a theoretical ED^2 based on these single measurements of illuminance. These calculated curves are drawn in Figs. 6 and 7 along with the experimental points. It is obvious that the theoretical curve fits the data from detector 2 (Fig. 7) very well while the curve obtained from the data of detector 1 (Fig. 6) does not. Here is a case where



Fig. 4: Plot of illuminance-distance squared product versus distance of a given flashlamp with detector 1.



Fig. 5: Plot of illuminance-distance squared product versus distance of a given flashlamp with detector 2.



Fig. 6: Data of Fig. 4 with theoretical curve derived from Eq. 25.



Fig. 7: Data of Fig. 5 with theoretical curve derived from Eq. 25.

intuition based on the inverse square approximation may lead to an incorrect interpretation of the experimental results.

One might ask how well the inverse square approximation with a term to correct the distance as in Eq. (17) would fit these two sets of data. The answer is given in Figs. 8 and 9. Here we have plotted the curves obtained by fitting all the data to a function of the form given in Eq. (17). The fit appears to be reasonably good for detector 1 and fair for detector 2. On the other hand, the distance correction term δ computed from the data of detector 1 is 0.23 cm and from detector 2 it is 2.18 cm--a difference of a factor of ten for two sets of data obtained using the same lamp. The inverse square approximation itself gives no indication of which detector is more linear. In fact, even after fitting all the data to an equation with a correction for distance one may still be misled into believing that detector 1 is better than detector 2. One must know in some detail how the illuminance from a source varies with distance before beginning the data analysis.

6. Curve-Fitting

If one were to interpolate or possibly to extrapolate to large distances in order to determine the so-called "constant of the source", that is the intensity, one must fit the experimental data to a smooth curve. Using Eq. (25) we may fit our data to several different approximations of the same function. We may take as the best approximation an equation of the form:

$$1 = \frac{A_1}{ED(D+d)} + \frac{A_2}{ED(D+d)^3} + \frac{A_3}{ED^3(D+d)} \qquad (26)$$

or we may assume that A_2 and A_3 will be very small and just fit to an equation of the form

$$1 = \frac{B_1}{ED(D+d)} \qquad (27)$$

In both Eqs. (26) and (27) we would have to use the approximate values of d. Since the lamp is placed behind a filter of unknown index of refraction and since it was difficult to measure the exact separation between the lamp and the aperture, it may be well to include d in the unknown coefficients to be determined from the fitting. Equation (27) can be rewritten as

$$ED^2 + dED = B_1 \qquad (28a)$$

or

$$1 = C_1ED^2 + C_2ED$$

Taking the data obtained from detector 2 we used a least square method to fit them to each of the above three equations. These curves are plotted in Fig. 10 along with the original data. All three equations provide a reasonable fit to the data with Eqs. (26) and (28) having a better fit. The difference between interpolated values of $\bar{E}$ at 0.6 m are: from Eq. (26), 1.774 lux; from Eq. (27), 1.764 lux; and Eq. (28), 1.774 lux. The extrapolated values at 0.4 m are: from Eq. (26), 3.920 lux; from Eq. (27), 3.849 lux; and from Eq. (28), 3.900 lux. The "intensity" computed from each of these equations, that is, the value of ED^2 at 10 m is: from Eq. (26), 0.671 cd; from Eq. (27), 0.675 cd; and from Eq. (28), 0.668 cd. The difference between the interpolations obtained using Eqs. (26) and (28) are negligible, however, the differences between both extrapolations are about 0.5%.

The last extrapolation to very large distances shows that the exact form of the equation to which one chooses to fit the data affects the determination of the "intensity". We have



Fig. 8: Data of Fig. 4 with curve fitting of Eq. 17.



Fig. 9: Data of Fig. 5 with curve fitting of Eq. 17.



Fig. 10: Data of Fig. 5 with three curve fitting equations. __.__. is Eq. 27. __ __ __ is Eq. 28. _____ is Eq. 26.

not attempted to weigh any of our data. Had we done so, to account for the increase in noise at large distances for example, this would have affected our determination of the "intensity" even more.

This exercise in data analysis and curve fitting alludes to a serious problem in photometry: having intensity as the base unit and fundamental quantity. This is unfortunate because, as demonstrated in this work, the accurate determination of the intensity of a source is subject to unquantifiable errors which are of necessity introduced by curve-fitting.

7. Summary

We have presented a method which is useful in determining the energy irradiance or illuminance output of single flashes from a pulsed source. We have described a general approach, discussed a specific technique employed, and presented in detail a general model which can be used to describe the illuminance (and irradiance) as a function of distance from an aperture-lamp configuration typical of flash sources. We have examined a number of experimental details and the effects they have upon the results of these measurements. Finally, we have indicated some of the difficulties encountered in curve fitting, and how they could yield erroneous results. Although several of the techniques employed in this measurement could obviously be refined and improved if desired, the basic approach should be satisfactory to many needs.

****

The authors gratefully acknowledge the contributions and stimulating discussions with J. Geist, particularly concerning the models for data analysis.

8. References

1. Geist, J. and Zalewski, E. F., Applied Optics 12, 435 (February 1973).
2. C.I.E. Publication No. 17: International Lighting Vocabulary. Third edition. Paris: Bureau Central de la C.I.E. 1970.
3. Margenau, H. and Murphy, G., The Mathematics of Physics and Chemistry, (D. Van Nostrand Company, New Jersey) p 477, (1956).
4. Zalewski, E. F.; Schaefer, A. R.; Mohan, K.; and McSparron, D. A.; Optical Radiation Measurements: Photometric Instrumentation and Research (1970 to 1971), NBS Technical Note 594-2, (1972).

TABLE 1

Illuminance of a Given Flashlamp as a Function of Distance as Measured by Detector 1.

D Distance (m)	E Illuminance (lux)	ED^2
.4141	3.6841	.63174
	3.6828	.63152
	3.6815	.63130
.5141	2.4277	.64164
	2.4316	.64167
	2.4251	.64095
.6141	1.6843	.63518
	1.6856	.63567
	1.6856	.63567
.7141	1.2519	.63839
	1.2489	.63686
	1.2508	.63783
.8141	.96126	.63708
	.96218	.63769
	.96336	.63847
.9141	.76875	.64235
	.76757	.64136
	.76678	.64070
1.0141	.62282	.64051
	.62296	.64065
	.62125	.63889
1.1141	.52048	.64603
	.51891	.64408
	.52101	.64669
1.2141	.43672	.64374
	.43685	.64393
	.43738	.64471
1.4141	.32103	.64196
	.32129	.64248
	.31985	.63960
1.5141	.28125	.64476
	.28229	.64715
	.28007	.64206
1.6141	.24526	.63898
	.24617	.64135
	.24591	.64067
1.7141	.22065	.64830
	.21830	.64139
	.21725	.63831
1.8141	.19552	.64345
	.19382	.63785
	.19395	.63828
1.9141	.17537	.64252
	.17458	.63962
	.17485	.64061

Table 1 (continued)

D Distance (m)	E Illuminance (lux)	ED^2
2.0141	.15796	.64078
	.15914	.64557
	.15940	.64662
2.1141	.13912	.62178
	.14082	.62938
	.14003	.62585
2.2141	.13218	.64798
	.13179	.64607
	.13362	.65504
2.3141	.12197	.65316
	.11988	.64196
	.11883	.63634
2.3641	.11504	.64295
	.11517	.64368
	.11438	.63927

TABLE 2

Illuminance of a Given Flashlamp as a Function of Distance as Measured by Detector 2.

D Distance (m)	E Illuminance (lux)	ED^2
.5152	2.3896	.63427
	2.3861	.63334
	2.3922	.63496
	2.3913	.63472
	2.3887	.63403
.7652	1.1017	.64508
	1.1035	.64613
	1.1026	.64561
	1.1009	.64461
	1.1026	.64561
1.0152	.63157	.65091
	.63243	.65180
	.63243	.65180
	.63139	.65073
	.63209	.65145
1.2152	.44104	.65129
	.44174	.65232
	.44070	.65079
	.44209	.65284
	.44287	.65400
1.5152	.28609	.65681
	.28626	.65720
	.28635	.65741
	.28617	.65700
1.7652	.21235	.66167
	.21313	.66410
	.21217	.66111
	.21226	.66139
	.21200	.66058

U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET	1. PUBLICATION OR REPORT NO. NBS TN-935	2. Gov't Accession No.	3. Recipient's Accession No.

4. TITLE AND SUBTITLE

MEASUREMENT OF ENERGY IRRADIANCE FROM SINGLE PULSE SOURCES

5. Publication Date: February 1977

6. Performing Organization Code

7. AUTHOR(S)

A. R. Schaefer and E. F. Zalewski

8. Performing Organ. Report No.

9. PERFORMING ORGANIZATION NAME AND ADDRESS

NATIONAL BUREAU OF STANDARDS
DEPARTMENT OF COMMERCE
WASHINGTON, D.C. 20234

10. Project/Task/Work Unit No.

11. Contract/Grant No.

12. Sponsoring Organization Name and Complete Address *(Street, City, State, ZIP)*

13. Type of Report & Period Covered: Final

14. Sponsoring Agency Code

15. SUPPLEMENTARY NOTES

16. ABSTRACT *(A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here.)*

A method of measuring the energy irradiance from a single pulsed source, such as a xenon flashtube, is presented. Details of the approach used in this particular measurement are given, along with a discussion of the data analysis. In particular, we consider some of the inherent problems encountered in curve fitting which are also common to other radiometric and photometric measurements.

17. KEY WORDS *(six to twelve entries; alphabetical order; capitalize only the first letter of the first key word unless a proper name; separated by semicolons)* Curve fitting; energy irradiance; flashtube; pulsed source; radiometry.

18. AVAILABILITY

[x] Unlimited

[ ] For Official Distribution. Do Not Release to NTIS

[X] Order From Sup. of Doc., U.S. Government Printing Office Washington, D.C. 20402, SD Cat. No. C13.46:935

[ ] Order From National Technical Information Service (NTIS) Springfield, Virginia 22151

19. SECURITY CLASS (THIS REPORT): UNCLASSIFIED

20. SECURITY CLASS (THIS PAGE): UNCLASSIFIED

21. NO. OF PAGES: 32

22. Price $.75

USCOMM-DC 29042-P74

BACK BY POPULAR DEMAND

MORE ENERGY AND DOLLAR SAVING IDEAS FROM EPIC

for you, the cost-conscious manager, in a new supplement to EPIC.

This first EPIC Supplement builds on the success of the original *Energy Conservation Program Guide for Industry and Commerce,* a comprehensive handbook that has helped thousands of managers establish or improve energy conservation programs in their organizations.

In the 89-page EPIC Supplement, you will find:

- A reorganized and simplified explanation of how to implement an energy conservation program.
- More than 50 additional Energy Conservation Opportunities to save energy in 13 different functional and operating areas.
- Two dozen new case studies that let you estimate your potential for conserving energy and reducing costs.
- New sources of assistance.

Industrial and commercial firms, hospitals, universities, research institutes, public utilities—ANY organization—whatever its size—that uses energy and has to pay the bills will want a copy of the EPIC Supplement.

The EPIC Supplement is intended to be used with the original EPIC guide. If you already have a copy of EPIC, don't delay in ordering the Supplement. Make EPIC part of your continuing plan to control energy use and costs.

To order your copy of EPIC Supplement, send $2.25 per copy (check, money order, or Superintendent of Documents coupons) to Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402. Use S.D. Catalog Number C13.11:115/1. To order EPIC, send $2.90 a copy to Superintendent of Documents, S.D. Catalog Number C13.11:115. Discount of 25% on orders of 100 copies or more mailed to one address.

U.S. DEPARTMENT OF COMMERCE/National Bureau of Standards
FEDERAL ENERGY ADMINISTRATION/Energy Conservation and Environment

SINGLE CRYSTAL DATA

REVISED! UPDATED!

In 1954, the first edition of CRYSTAL DATA (Determinative Tables and Systematic Tables) was published as Memoir 60 of the Geological Society of America. In 1960, the second edition of the Determinative Tables was issued as Monograph 5 of the American Crystallographic Association, and in 1967, the Systematic Tables were issued as Monograph 6. These editions proved extremely valuable to crystallographers throughout the world. Recognizing the need for updated crystallographic information, the National Bureau of Standards Office of Standard Reference Data has sponsored the issuance of a new edition.

This, the THIRD EDITION, should be of particular interest not only to crystallographers but also to chemists, mineralogists, physicists and individuals in related fields of study. The current edition, which comprises two volumes, Organic and Inorganic, is a thoroughly revised and updated work, containing over 25,000 entries.

The entries are listed, *within each crystal system*, according to increasing values of a determinative number: a/*b* ratio in trimetric systems, c/a ratio in dimetric systems, and cubic cell edge *a*, in the isometric system. In addition, the following information is given:



INORGANIC VOLUME $50.00

ORGANIC VOLUME $30.00

Plus shipping and handling

axial ratio(s) and interaxial angles not fixed by symmetry, cell dimensions, space group or diffraction aspect, number of formula units per unit cell, crystal structure, (whether determined), measured density and x-ray calculated density. Also listed is the name of the compound and synonym(s), chemical formula, literature reference and transformation matrix. When available, the crystal structure type, crystal habit, cleavages, twinning, color, optical properties, indices of refraction, optical orientation, melting point and transition point are also listed.

THIS EDITION culminates years of effort by J. D. H. Donnay, Johns Hopkins University, Helen M. Ondik, National Bureau of Standards, Sten Samson, California Institute of Technology, Quintin Johnson, Lawrence Radiation Laboratory, Melvin H. Mueller, Argonne National Laboratory, Gerard M. Wolten, Aerospace Corporation, Mary E. Mrose, U.S. Geological Survey, Olga Kennard and David G. Watson, Cambridge University, England and Murray Vernon King, Massachusetts General Hospital.

Shipments are made via insured parcel post. Additional charges for shipments by air or commercial carrier.
TERMS: Domestic—30 days Foreign—prepayment required. Address all orders to:
JOINT COMMITTEE ON POWDER DIFFRACTION STANDARDS 1601 Park Lane, Swarthmore, Pennsylvania 19081

Please accept my order for CRYSTAL DATA, DETERMINATIVE TABLES, Third Edition, Donnay/Ondik.

☐ **Organic Volume**

☐ **Inorganic Volume**

Ship to: _____

_____

Signature _____

NATIONAL STANDARD NSRDS REFERENCE DATA SYSTEM

JCPDS

MORE ENERGY AND DOLLAR SAVING IDEAS FROM EPIC

for you, the cost-conscious manager, in a new supplement to EPIC.

This first EPIC Supplement builds on the success of the original *Energy Conservation Program Guide for Industry and Commerce*, a comprehensive handbook that has helped thousands of managers establish or improve energy conservation programs in their organizations.

In the EPIC Supplement, you will find:

- A reorganized and simplified explanation of how to implement an energy conservation program.
- More than 50 additional Energy Conservation Opportunities to save energy in 13 different functional and operating areas.
- Two dozen new case studies that let you estimate your potential for conserving energy and reducing costs.
- New sources of assistance.

Any organization—whatever its size—that uses energy and has to pay the bills, will want a copy of the EPIC supplement.

The EPIC Supplement is intended to be used with the original EPIC guide. If you already have a copy of EPIC, don't delay in ordering the Supplement.

To order your copy of EPIC Supplement, send $2.25 per copy (check, money order, or Superintendent of Documents coupons) to Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402. Use S.D. Catalog Number C13.11:115/1. To order EPIC, send $2.90 a copy to Superintendent of Documents, S.D. Catalog Number C13.11:115. Discount of 25% on orders of 100 copies or more mailed to one address.

U.S. DEPARTMENT OF COMMERCE/National Bureau of Standards
FEDERAL ENERGY ADMINISTRATION/Energy Conservation and Environment

☆ U. S. GOVERNMENT PRINTING OFFICE :1977--240-848/25

NBS TECHNICAL PUBLICATIONS

PERIODICALS

JOURNAL OF RESEARCH reports National Bureau of Standards research and development in physics, mathematics, and chemistry. It is published in two sections, available separately:

• **Physics and Chemistry (Section A)**
Papers of interest primarily to scientists working in these fields. This section covers a broad range of physical and chemical research, with major emphasis on standards of physical measurement, fundamental constants, and properties of matter. Issued six times a year. Annual subscription: Domestic, $17.00; Foreign, $21.25.

• **Mathematical Sciences (Section B)**
Studies and compilations designed mainly for the mathematician and theoretical physicist. Topics in mathematical statistics, theory of experiment design, numerical analysis, theoretical physics and chemistry, logical design and programming of computers and computer systems. Short numerical tables. Issued quarterly. Annual subscription: Domestic, $9.00; Foreign, $11.25.

DIMENSIONS/NBS (formerly Technical News Bulletin)—This monthly magazine is published to inform scientists, engineers, businessmen, industry, teachers, students, and consumers of the latest advances in science and technology, with primary emphasis on the work at NBS. The magazine highlights and reviews such issues as energy research, fire protection, building technology, metric conversion, pollution abatement, health and safety, and consumer product performance. In addition, it reports the results of Bureau programs in measurement standards and techniques, properties of matter and materials, engineering standards and services, instrumentation, and automatic data processing.

Annual subscription: Domestic, $9.45; Foreign, $11.85.

NONPERIODICALS

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a world-wide program coordinated by NBS. Program under authority of National Standard Data Act (Public Law 90-396).

NOTE: At present the principal publication outlet for these data is the Journal of Physical and Chemical Reference Data (JPCRD) published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements available from ACS, 1155 Sixteenth St. N.W., Wash. D. C. 20056.

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The purpose of the standards is to establish nationally recognized requirements for products, and to provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order **above** *NBS publications from: Superintendent of Documents, Government Printing Office, Washington, D.C. 20402.*

Order **following** *NBS publications—NBSIR's and FIPS from the National Technical Information Services, Springfield, Va. 22161.*

Federal Information Processing Standards Publications (FIPS PUBS)—Publications in this series collectively constitute the Federal Information Processing Standards Register. Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

NBS Interagency Reports (NBSIR)—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Services (Springfield, Va. 22161) in paper copy or microfiche form.

BIBLIOGRAPHIC SUBSCRIPTION SERVICES

The following current-awareness and literature-survey bibliographies are issued periodically by the Bureau:

Cryogenic Data Center Current Awareness Service. A literature survey issued biweekly. Annual subscription: Domestic, $20.00; Foreign, $25.00.

Liquified Natural Gas. A literature survey issued quarterly. Annual subscription: $20.00.

Superconducting Devices and Materials. A literature survey issued quarterly. Annual subscription: $20.00. Send subscription orders and remittances for the preceding bibliographic services to National Bureau of Standards, Cryogenic Data Center (275.02) Boulder, Colorado 80302.

U.S. DEPARTMENT OF COMMERCE
National Bureau of Standards
Washington, D.C. 20234

OFFICIAL BUSINESS

Penalty for Private Use, $300

POSTAGE AND FEES PAID
U.S. DEPARTMENT OF COMMERCE
COM-215

U.S.MAIL

SPECIAL FOURTH-CLASS RATE.
BOOK